

[برگشت \(/IACLD/Panel/Labs/Answers/34407\)](#)

دوره چهل و پنجم تابستان 1402

S1.1 Results

نمونه مجهول ایمونوسرولوژی برای آزمایش HBsAg, HIVAb, S1.1 HCVAb

کد نمونه 2560	EQAP-10814	آزمایشگاه نوبل
---------------	------------	----------------

S1/1 HBs Ag	S1/1 HCV Ab	S1/1 HIV Ab
-------------	-------------	-------------

S1/1 HBs Ag

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان بر اساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ **+** بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت **←** مشخص شده است.

- ثبت نام کنندگان 1677
- شرکت کنندگان 1588
- کل درصد مشارکت 94.69 %
- پاسخهای دریافتی 1556
- درصد پاسخدهی 97.98 %

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	111	108	0	3	%97.30	%0.00	%2.70
ECL	134	130	2	2	%97.01	%1.49	%1.49

ELFA	96	93	0	3	%96.88	%0.00	%3.13
ELISA	1189	1110	18	61	%93.36	%1.51	%5.13
Immunochromatography	9	2	3	4	%22.22	%33.33	%44.44
Other	11	7	1	3	%63.64	%9.09	%27.27

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Not Reported	1	
CL	Not Reported	Positive	2	
CL	Abbott	Not Reported	1	
CL	Abbott	Positive	27	←
CL	Architect	Positive	7	
CL	Biomerieux	Positive	1	
CL	Dia Sorin	Positive	19	
CL	Liaison	Positive	8	
CL	Mindray	Negative	2	
CL	Mindray	Positive	15	
CL	Ortho	Positive	3	
CL	Pishtaz Teb	Positive	1	
CL	Roche	Negative	1	
CL	Siemens	Positive	25	
ECL	Not Reported	Positive	3	
ECL	Abbott	Negative	1	

Method	Kit	Result	Count	Your Answer
ECL	Abbott	Positive	3	
ECL	Architect	Positive	2	
ECL	Biomerieux	Positive	6	
ECL	Cobas	Positive	7	
ECL	Dia Sorin	Positive	6	
ECL	ImmunoLab	Positive	1	
ECL	Liaison	Positive	4	
ECL	Mindray	Positive	4	
ECL	Pishtaz Teb	Positive	2	
ECL	Roche	Indeterminate	2	
ECL	Roche	Negative	1	
ECL	Roche	Positive	84	
ECL	Siemens	Positive	7	
ECL	Tosoh	Positive	1	
ELFA	Not Reported	Positive	5	
ELFA	Biomerieux	Not Reported	1	
ELFA	Biomerieux	Negative	3	
ELFA	Biomerieux	Positive	87	
ELFA	Siemens	Positive	1	
ELISA	Not Reported	Indeterminate	1	
ELISA	Not Reported	Negative	4	

Method	Kit	Result	Count	Your Answer
ELISA	Not Reported	Positive	36	
ELISA	Abbott	Positive	1	
ELISA	Abon	Positive	1	
ELISA	Ai De	Positive	2	
ELISA	Anti Human	Positive	1	
ELISA	Arya Mabna Tashkhis	Indeterminate	1	
ELISA	Arya Mabna Tashkhis	Positive	44	
ELISA	Bio Kit	Positive	1	
ELISA	Bio-Pharm	Positive	3	
ELISA	Biorad	Positive	1	
ELISA	Diapro	Positive	18	
ELISA	Euroimmun	Positive	1	
ELISA	Gbc	Indeterminate	5	
ELISA	Gbc	Negative	4	
ELISA	Gbc	Positive	49	
ELISA	General Biologicals(GB)	Not Reported	1	
ELISA	General Biologicals(GB)	Negative	3	
ELISA	General Biologicals(GB)	Positive	40	
ELISA	Mindray	Positive	1	
ELISA	Monobind	Positive	3	
ELISA	Others	Positive	1	

Method	Kit	Result	Count	Your Answer
ELISA	Padtan Teb	Positive	5	
ELISA	Pishtaz Teb	Not Reported	2	
ELISA	Pishtaz Teb	Indeterminate	9	
ELISA	Pishtaz Teb	Negative	44	
ELISA	Pishtaz Teb	Positive	723	
ELISA	Sam Fan Avaran	Positive	1	
ELISA	Siemens	Positive	1	
ELISA	Surase B-96(Tmb)	Indeterminate	1	
ELISA	Surase B-96(Tmb)	Positive	2	
ELISA	Wantai	Indeterminate	1	
ELISA	Wantai	Negative	6	
ELISA	Wantai	Positive	175	
Immunochromatography	Abon	Indeterminate	2	
Immunochromatography	Abon	Negative	4	
Immunochromatography	Arya Mabna Tashkhis	Positive	1	
Immunochromatography	Medical Biological Service	Indeterminate	1	
Immunochromatography	Mindray	Positive	1	
Other	Not Reported	Positive	3	
Other	Abbott	Positive	1	
Other	Abon	Negative	2	
Other	Acon	Negative	1	

Method	Kit	Result	Count	Your Answer
Other	Biomerieux	Positive	2	
Other	Dia Sorin	Positive	1	
Other	Siemens	Indeterminate	1	
RIA	Abon	Positive	2	
RIA	Biomerieux	Positive	3	

Copyright © 2023 IACLD. All rights reserved.