

[برگشت \(/IACLD/Panel/Labs/Answers/34407\)](#)

دوره چهل و پنجم تابستان 1402

S2 Results

نمونه مجهول ToRCH

کد نمونه 1205	EQAP-10814	آزمایشگاه نوبل
---------------	------------	----------------

Anti-CMV (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

Anti-Rubella (IgG)

Anti-Toxoplasma (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان بر اساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ **+** بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت **←** مشخص شده است.

- ثبت نام کنندگان 222
- شرکت کنندگان 198
- کل درصد مشارکت 89.19 %
- پاسخهای دریافتی 82
- درصد پاسخدهی 41.41 %

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	18	18	0	0	%100.00	%0.00	%0.00
ECL	4	4	0	0	%100.00	%0.00	%0.00

ELFA	5	4	1	0	%80.00	%20.00	%0.00
ELISA	48	45	0	3	%93.75	%0.00	%6.25
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00
Other	7	7	0	0	%100.00	%0.00	%0.00

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
CL	Chorus	Positive	1	
CL	Dia Sorin	Positive	10	
CL	Diesse-Tec	Positive	1	
CL	ImmunoDiagnostic Systems	Positive	1	
CL	Liaison	Positive	1	
CL	Others	Positive	3	
ECL	Cobas	Positive	1	
ECL	Dia Sorin	Positive	1	
ECL	Maglumi	Positive	1	
ECL	Roche	Positive	1	
ELFA	Chorus	Positive	2	
ELFA	Diesse-Tec	Indeterminate	1	
ELFA	Maglumi	Positive	1	
ELFA	Orgentec	Positive	1	
ELISA	Not Reported	Positive	4	

Method	Kit	Result	Count	Your Answer
ELISA	Chorus	Negative	1	
ELISA	Chorus	Positive	2	
ELISA	Diesse-Tec	Positive	1	
ELISA	Orgentec	Positive	10	
ELISA	Pishtaz Teb	Not Reported	1	
ELISA	Pishtaz Teb	Negative	1	
ELISA	Pishtaz Teb	Positive	10	
ELISA	Vircell	Negative	1	
ELISA	Vircell	Positive	18	
Other	Chorus	Positive	2	
Other	Diesse-Tec	Positive	1	
Other	Orgentec	Positive	4	