

انجمن دکترای علوم آزمایشگاهی ایران

• دکتر علی عجمی

نوبل

- منوی اصلی
- صفحه نخست
- دوره های ارزیابی خارجی
- خروج

دوره چهل و یکم زمستان 1400

S2 Results

نمونه مجهول ToRCH

آزمایشگاه نوبل	EQAP-10814	کد نمونه 1108
----------------	------------	---------------

- Anti-CMV (IgG)
- Anti-HSV1 (Anti-Herpes Type1 IgG)
- Anti-Rubella (IgG)
- Anti-Toxoplasma (IgG)

Anti-CMV (IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است .

- ثبت نام کنندگان 227
- شرکت کنندگان 199
- کل درصد مشارکت 87.67 %
- پاسخهای دریافتی 170
- درصد پاسخدهی 85.43 %

Method	Total	Positive	Indeterminate	Count			Percent
				Negative	Positive	Indeterminate	Negative
CL	57	57	0	0	%100.00	%0.00	%0.00
ECL	30	30	0	0	%100.00	%0.00	%0.00
ELFA	25	25	0	0	%100.00	%0.00	%0.00
ELISA	56	56	0	0	%100.00	%0.00	%0.00
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00
Other	2	2	0	0	%100.00	%0.00	%0.00

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	3	
CL	Abbott	Positive	5	
CL	Architect	Positive	2	
CL	Autobio Diagnostic	Not Reported	1	
CL	Autobio Diagnostic	Positive	2	←
CL	Biomerieux	Positive	1	
CL	Chorus	Positive	1	
CL	Dia Sorin	Positive	22	
CL	ImmunoLab	Positive	1	
CL	Liaison	Positive	10	
CL	Maglumi	Not Reported	1	
CL	Maglumi	Positive	1	
CL	Siemens	Not Reported	1	

Method	Kit	Result	Count	Your Answer
CL	Siemens	Positive	9	
ECL	Not Reported	Positive	1	
ECL	Abbott	Positive	2	
ECL	Architect	Positive	3	
ECL	Autobio Diagnostic	Positive	1	
ECL	Biomerieux	Positive	1	
ECL	Cobas	Positive	2	
ECL	Dia Sorin	Positive	1	
ECL	Liaison	Positive	1	
ECL	Monobind	Positive	1	
ECL	Roche	Not Reported	1	
ECL	Roche	Positive	16	
ECL	Siemens	Positive	1	
ELFA	Not Reported	Positive	1	
ELFA	Biomerieux	Positive	22	
ELFA	Chorus	Positive	1	
ELFA	Diesse-Tec	Positive	1	
ELISA	Not Reported	Positive	2	
ELISA	Chorus	Positive	5	
ELISA	Euroimmun	Positive	1	
ELISA	Pishtaz Teb	Positive	39	
ELISA	Trinitybiotech	Positive	1	
ELISA	Vircell	Positive	8	
Other	Not Reported	Positive	1	
Other	Pishtaz Teb	Positive	1	

Copyright © 2022 [IACLD](#). All rights reserved.

• انجمن دکترای علوم آزمایشگاهی ایران

• دکتر علی عجمی

نوبل

- منوی اصلی
- صفحه نخست
- دوره های ارزیابی خارجی
- خروج

دوره چهل و یکم زمستان 1400

S2 Results

نمونه مجهول ToRCH

آزمایشگاه نوبل	EQAP-10814	کد نمونه 1108
----------------	------------	---------------

- Anti-CMV (IgG)
- Anti-HSV1 (Anti-Herpes Type1 IgG)
- Anti-Rubella (IgG)
- Anti-Toxoplasma (IgG)

Anti-HSV1 (Anti-Herpes Type1 IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش

تذکر. نتیجه مورد قبول برای این آزمایش پاسخ  بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت  مشخص شده است .

- ثبت نام کنندگان 227
- شرکت کنندگان 199

- کل درصد مشارکت 87.67٪
- پاسخهای دریافتی 68
- درصد پاسخدهی 34.17٪

Method	Total	Count				Percent		
		Positive	Indeterminate	Negative	Positive	Indeterminate	Negative	
CL	21	20	1	0	%95.24	%4.76	%0.00	
ECL	4	3	0	1	%75.00	%0.00	%25.00	
ELFA	4	4	0	0	%100.00	%0.00	%0.00	
ELISA	36	33	1	2	%91.67	%2.78	%5.56	
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00	
Other	3	3	0	0	%100.00	%0.00	%0.00	

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Not Reported	2	
CL	Not Reported	Indeterminate	1	
CL	Not Reported	Positive	1	
CL	Asialion	Positive	1	
CL	Autobio Diagnostic	Positive	1	

Method	Kit	Result	Count	Your Answer
CL	Chorus	Positive	1	
CL	Dia Lab	Positive	1	
CL	Dia Sorin	Positive	9	
CL	Liaison	Positive	4	
CL	Maglumi	Positive	2	
ECL	Autobio Diagnostic	Positive	1	
ECL	Liaison	Positive	1	
ECL	Mindray	Negative	1	
ECL	Roche	Not Reported	1	
ECL	Roche	Positive	1	
ELFA	Not Reported	Not Reported	1	
ELFA	Chorus	Positive	2	
ELFA	Diesse-Tec	Positive	2	
ELISA	Not Reported	Positive	3	
ELISA	Chorus	Positive	7	
ELISA	Diesse-Tec	Positive	2	
ELISA	Euroimmun	Positive	2	
ELISA	Lion	Positive	1	
ELISA	Pishtaz Teb	Positive	7	
ELISA	Vircell	Indeterminate	1	
ELISA	Vircell	Negative	2	

Method	Kit	Result	Count	Your Answer
ELISA	Vircell	Positive	11	←
Other	Not Reported	Positive	1	
Other	Chorus	Positive	1	
Other	Orgentec	Positive	1	
Other	Pishtaz Teb	Not Reported	1	

Copyright © 2022 [IACLD](#). All rights reserved.

[EQAP Panel](#) Toggle navigation_

انجمن دکترای علوم آزمایشگاهی ایران

• دکتر علی عجمی

نوبل

- منوی اصلی
- صفحه نخست
- دوره های ارزیابی خارجی
- خروج

دوره چهل و یکم زمستان 1400

S2 Results

نمونه مجهول ToRCH

آزمایشگاه نوبل	EQAP-10814	کد نمونه 1108
----------------	------------	---------------

- Anti-CMV (IgG)
- Anti-HSV1 (Anti-Herpes Type1 IgG)
- Anti-Rubella (IgG)
- Anti-Toxoplasma (IgG)

Anti-Rubella (IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است .

- ثبت نام کنندگان 227
- شرکت کنندگان 199
- کل درصد مشارکت 87.67 %
- پاسخهای دریافتی 186
- درصد پاسخدهی 93.47 %

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	61	61	0	0	%100.00	%0.00	%0.00
ECL	29	29	0	0	%100.00	%0.00	%0.00
ELFA	16	15	0	1	%93.75	%0.00	%6.25
ELISA	76	73	0	3	%96.05	%0.00	%3.95
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00
Other	4	4	0	0	%100.00	%0.00	%0.00

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	2	
CL	Abbott	Positive	8	
CL	Architect	Positive	4	
CL	Autobio Diagnostic	Positive	4	←
CL	Biomerieux	Positive	1	

Method	Kit	Result	Count	Your Answer
CL	Chorus	Positive	1	
CL	Dia Sorin	Positive	22	
CL	Liaison	Positive	10	
CL	Maglumi	Not Reported	1	
CL	Roche	Positive	1	
CL	Siemens	Not Reported	1	
CL	Siemens	Positive	8	
ECL	Not Reported	Positive	1	
ECL	Abbott	Positive	2	
ECL	Architect	Positive	2	
ECL	Autobio Diagnostic	Positive	1	
ECL	Cobas	Positive	2	
ECL	Dia Sorin	Positive	1	
ECL	Roche	Not Reported	1	
ECL	Roche	Positive	20	
ELFA	Not Reported	Positive	1	
ELFA	Biomerieux	Positive	13	
ELFA	Diesse-Tec	Positive	1	
ELFA	Pishtaz Teb	Negative	1	
ELISA	Not Reported	Positive	1	
ELISA	Biomerieux	Positive	1	

Method	Kit	Result	Count	Your Answer
ELISA	Chorus	Positive	3	
ELISA	Diesse-Tec	Positive	1	
ELISA	Euroimmun	Positive	2	
ELISA	Padtan Teb	Positive	1	
ELISA	Pishtaz Teb	Negative	3	
ELISA	Pishtaz Teb	Positive	57	
ELISA	Trinitybiotech	Positive	1	
ELISA	Vircell	Positive	6	
Other	Not Reported	Positive	1	
Other	Pishtaz Teb	Positive	1	
Other	Siemens	Positive	2	

Copyright © 2022 [IACLD](#). All rights reserved.

انجمن دکترای علوم آزمایشگاهی ایران

• دکتر علی عجمی

نوبل

- منوی اصلی
- صفحه نخست
- دوره های ارزیابی خارجی
- خروج

دوره چهل و یکم زمستان 1400

S2 Results

نمونه مجهول ToRCH

آزمایشگاه نوبل	EQAP-10814	کد نمونه 1108
----------------	------------	---------------

- Anti-CMV (IgG)
- Anti-HSV1 (Anti-Herpes Type1 IgG)
- Anti-Rubella (IgG)
- Anti-Toxoplasma (IgG)

Anti-Toxoplasma (IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است .

- ثبت نام کنندگان 227
- شرکت کنندگان 199
- کل درصد مشارکت 87.67 %
- پاسخهای دریافتی 179
- درصد پاسخدهی 89.95 %

Method	Total	Positive	Indeterminate	Negative	Count		Percent	
					Positive	Indeterminate	Negative	Negative
CL	59	23	11	25	%38.98	%18.64	%42.37	
ECL	28	21	2	5	%75.00	%7.14	%17.86	
ELFA	16	3	11	2	%18.75	%68.75	%12.50	
ELISA	73	56	5	12	%76.71	%6.85	%16.44	
Immunochromatography	1	0	0	1	%0.00	%0.00	%100.00	
Other	2	2	0	0	%100.00	%0.00	%0.00	

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Indeterminate	1	
CL	Not Reported	Positive	1	
CL	Abbott	Indeterminate	1	
CL	Abbott	Positive	4	
CL	Architect	Indeterminate	2	
CL	Autobio Diagnostic	Positive	4	←
CL	Chorus	Indeterminate	1	
CL	Dia Sorin	Indeterminate	6	

Method	Kit	Result	Count	Your Answer
CL	Dia Sorin	Negative	13	
CL	Dia Sorin	Positive	2	
CL	Liaison	Negative	7	
CL	Liaison	Positive	3	
CL	Maglumi	Not Reported	1	
CL	Maglumi	Negative	2	
CL	Siemens	Not Reported	1	
CL	Siemens	Negative	3	
CL	Siemens	Positive	9	
ECL	Not Reported	Positive	1	
ECL	Abbott	Indeterminate	1	
ECL	Abbott	Positive	1	
ECL	Architect	Indeterminate	1	
ECL	Architect	Positive	2	
ECL	Cobas	Positive	2	
ECL	Dia Sorin	Negative	1	
ECL	Liaison	Negative	1	
ECL	Maglumi	Negative	1	
ECL	Monobind	Negative	1	
ECL	Roche	Not Reported	1	
ECL	Roche	Positive	15	

Method	Kit	Result	Count	Your Answer
ECL	Siemens	Negative	1	
ELFA	Biomerieux	Indeterminate	11	
ELFA	Biomerieux	Negative	1	
ELFA	Biomerieux	Positive	3	
ELFA	Diesse-Tec	Negative	1	
ELISA	Chorus	Indeterminate	1	
ELISA	Chorus	Negative	5	
ELISA	Euroimmun	Positive	1	
ELISA	ImmunoLab	Positive	1	
ELISA	Pishtaz Teb	Not Reported	1	
ELISA	Pishtaz Teb	Indeterminate	4	
ELISA	Pishtaz Teb	Negative	6	
ELISA	Pishtaz Teb	Positive	47	
ELISA	Trinitybiotech	Positive	1	
ELISA	Vircell	Negative	1	
ELISA	Vircell	Positive	6	
Immunochromatography	Maglumi	Negative	1	
Other	Not Reported	Positive	1	
Other	Pishtaz Teb	Positive	1	