

## • انجمن دکترای علوم آزمایشگاهی ایران

• دکتر علی عجمی

نوبل

- منوی اصلی
- صفحه نخست
- دوره های ارزیابی خارجی
- خروج

دوره سی و نهم تابستان 1400

### S1.2 Results

نمونه مجهول ایمنوسرولوژی برای آزمایش S1/2 HBsAg, HCVAb ,HIVAb

آزمایشگاه نوبل	EQAP-10814	1118 کد نمونه
----------------	------------	---------------

- [S1/2 HBs Ag](#)
- [S1/2 HCV Ab](#)
- [S1/2 HIV Ab](#)

### S1/2 HBs Ag

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ مثبت بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت مشخص شده است .

- ثبت نام کنندگان 1620
- شرکت کنندگان 1513
- کل درصد مشارکت 93.40 %
- پاسخهای دریافتی 1473

• درصد پاسخدهی 97.36٪

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	90	89	0	1	%98.89	%0.00	%1.11
ECL	133	129	2	2	%96.99	%1.50	%1.50
ELFA	129	126	0	3	%97.67	%0.00	%2.33
ELISA	1089	1028	16	45	%94.40	%1.47	%4.13
Immunochromatography	17	3	1	13	%17.65	%5.88	%76.47
Other	7	4	1	2	%57.14	%14.29	%28.57

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	4	
CL	Abbott	Positive	20	←
CL	Architect	Positive	4	
CL	Biomerieux	Positive	3	
CL	Dia Sorin	Positive	18	
CL	Liaison	Positive	11	
CL	Mindray	Not Reported	1	
CL	Mindray	Negative	1	
CL	Mindray	Positive	11	
CL	Others	Positive	1	
CL	Siemens	Positive	17	
ECL	Not Reported	Positive	4	

Method	Kit	Result	Count	Your Answer
ECL	Abbott	Positive	10	
ECL	Architect	Positive	2	
ECL	Biomerieux	Positive	3	
ECL	Cobas	Negative	1	
ECL	Cobas	Positive	6	
ECL	Dia Sorin	Positive	2	
ECL	Liaison	Positive	4	
ECL	Mindray	Positive	1	
ECL	Roche	Indeterminate	2	
ECL	Roche	Positive	95	
ECL	Siemens	Positive	2	
ECL	Tosoh	Negative	1	
ELFA	Not Reported	Positive	2	
ELFA	Abbott	Positive	1	
ELFA	Biomerieux	Negative	3	
ELFA	Biomerieux	Positive	120	
ELFA	Biorex	Positive	2	
ELFA	General Biologicals(GB)	Positive	1	
ELISA	Not Reported	Negative	3	
ELISA	Not Reported	Positive	43	
ELISA	Abon	Positive	1	
ELISA	Ai De	Positive	2	
ELISA	Apdia	Positive	1	
ELISA	Arya Mabna Tashkhis	Indeterminate	1	
ELISA	Arya Mabna Tashkhis	Negative	2	
ELISA	Arya Mabna Tashkhis	Positive	38	
ELISA	Autobio Diagnostic	Indeterminate	1	
ELISA	Autobio Diagnostic	Negative	1	
ELISA	Autobio Diagnostic	Positive	14	

Method	Kit	Result	Count	Your Answer
ELISA	Bio-Pharm	Negative	1	
ELISA	Bio-Pharm	Positive	8	
ELISA	Biorad	Positive	3	
ELISA	Biosystems	Positive	1	
ELISA	Diapro	Indeterminate	1	
ELISA	Diapro	Positive	18	
ELISA	Gbc	Negative	2	
ELISA	Gbc	Positive	17	
ELISA	General Biologicals(GB)	Not Reported	2	
ELISA	General Biologicals(GB)	Indeterminate	1	
ELISA	General Biologicals(GB)	Negative	1	
ELISA	General Biologicals(GB)	Positive	58	
ELISA	Mindray	Positive	2	
ELISA	Monobind	Positive	1	
ELISA	Others	Negative	1	
ELISA	Others	Positive	1	
ELISA	Padtan Teb	Not Reported	1	
ELISA	Padtan Teb	Positive	6	
ELISA	Pars Azmoon	Positive	1	
ELISA	Pishtaz Teb	Not Reported	2	
ELISA	Pishtaz Teb	Indeterminate	8	
ELISA	Pishtaz Teb	Negative	29	
ELISA	Pishtaz Teb	Positive	622	
ELISA	Spinreact	Negative	1	
ELISA	Spinreact	Positive	1	
ELISA	Surase B-96(Tmb)	Positive	6	
ELISA	Wantai	Not Reported	2	
ELISA	Wantai	Indeterminate	4	
ELISA	Wantai	Negative	4	

Method	Kit	Result	Count	Your Answer
ELISA	Wantai	Positive	184	
Immunochromatography	Not Reported	Negative	3	
Immunochromatography	Not Reported	Positive	1	
Immunochromatography	Abon	Negative	6	
Immunochromatography	Abon	Positive	2	
Immunochromatography	Addcare	Negative	2	
Immunochromatography	Diagnostic	Negative	1	
Immunochromatography	General Biologicals(GB)	Indeterminate	1	
Immunochromatography	Others	Negative	1	
Other	Abbott	Not Reported	1	
Other	Abbott	Indeterminate	1	
Other	Acon	Negative	1	
Other	Biomerieux	Not Reported	1	
Other	Biomerieux	Negative	1	
Other	Biomerieux	Positive	2	
Other	Siemens	Positive	2	
RIA	Not Reported	Positive	1	
RIA	Abon	Positive	1	
RIA	Pishtaz Teb	Positive	1	

Copyright © 2021 [IACLD](#). All rights reserved.

[EQAP Panel Toggle navigation](#)

• انجمن دکترای علوم آزمایشگاهی ایران

• دکتر علی عجمی

نوبل

- منوی اصلی
- صفحه نخست
- دوره های ارزیابی خارجی
- خروج

دوره سی و نهم تابستان 1400

## S1.2 Results

نمونه مجهول ایمنوسروولوژی برای آزمایش S1/2 HBsAg, HCVAb ,HIVAb

1118 کد نمونه EQAP-10814 آزمایشگاه نوبل

- [S1/2 HBs Ag](#)
- [S1/2 HCV Ab](#)
- [S1/2 HIV Ab](#)

## S1/2 HCV Ab

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ منفی بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت مشخص شده است .

- ثبت نام کنندگان 1620
- شرکت کنندگان 1513
- کل درصد مشارکت 93.40 %
- پاسخهای دریافتی 1398
- درصد پاسخدهی 92.40 %

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	35	2	0	33	%5.71	%0.00	%94.29
ECL	117	1	0	116	%0.85	%0.00	%99.15
ELFA	150	1	0	149	%0.67	%0.00	%99.33
ELISA	1067	33	5	1029	%3.09	%0.47	%96.44
Immunochromatography	18	0	0	18	%0.00	%0.00	%100.00

Other	6	0	0	6	%0.00	%0.00	%100.00
-------	---	---	---	---	-------	-------	---------

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Negative	1	
CL	Abbott	Negative	18	←
CL	Abbott	Positive	1	
CL	Architect	Negative	2	
CL	Biomerieux	Negative	4	
CL	Mindray	Negative	1	
CL	Others	Negative	1	
CL	Siemens	Negative	6	
CL	Siemens	Positive	1	
ECL	Not Reported	Negative	3	
ECL	Abbott	Negative	10	
ECL	Architect	Negative	2	
ECL	Biomerieux	Negative	5	
ECL	Cobas	Negative	9	
ECL	Mindray	Negative	1	
ECL	Roche	Negative	86	
ECL	Roche	Positive	1	
ELFA	Not Reported	Negative	2	
ELFA	Abbott	Negative	1	
ELFA	Architect	Negative	1	
ELFA	Bio Test	Negative	1	
ELFA	Biomerieux	Negative	141	
ELFA	General Biologicals(GB)	Negative	1	
ELFA	General Biologicals(GB)	Positive	1	
ELFA	Siemens	Negative	2	
ELISA	Not Reported	Indeterminate	1	
ELISA	Not Reported	Negative	55	

Method	Kit	Result	Count	Your Answer
ELISA	Not Reported	Positive	3	
ELISA	Acon	Negative	1	
ELISA	Apdia	Negative	1	
ELISA	Architect	Negative	1	
ELISA	Arya Mabna Tashkhis	Negative	23	
ELISA	Arya Mabna Tashkhis	Positive	5	
ELISA	Autobio Diagnostic	Negative	1	
ELISA	Bio Kit	Negative	1	
ELISA	Bio-Pharm	Negative	5	
ELISA	Biorad	Negative	3	
ELISA	Dia Sorin	Negative	4	
ELISA	Diapro	Negative	22	
ELISA	Diapro	Positive	1	
ELISA	Euroimmun	Negative	1	
ELISA	Gbc	Negative	57	
ELISA	General Biologicals(GB)	Not Reported	2	
ELISA	General Biologicals(GB)	Indeterminate	1	
ELISA	General Biologicals(GB)	Negative	117	
ELISA	General Biologicals(GB)	Positive	1	
ELISA	Mindray	Negative	2	
ELISA	Murex	Negative	2	
ELISA	Non Base	Negative	5	
ELISA	Padtan Teb	Indeterminate	1	
ELISA	Padtan Teb	Negative	4	
ELISA	Pars Azmoon	Negative	2	
ELISA	Pishtaz Teb	Not Reported	2	
ELISA	Pishtaz Teb	Indeterminate	2	
ELISA	Pishtaz Teb	Negative	581	
ELISA	Pishtaz Teb	Positive	20	



Method	Kit	Result	Count	Your Answer
ELISA	Spinreact	Negative	1	
ELISA	Wantai	Not Reported	1	
ELISA	Wantai	Negative	140	
ELISA	Wantai	Positive	3	
Immunochromatography	Not Reported	Negative	3	
Immunochromatography	Abon	Negative	9	
Immunochromatography	Acon	Negative	1	
Immunochromatography	Addcare	Negative	2	
Immunochromatography	Diagnostic	Negative	1	
Immunochromatography	General Biologicals(GB)	Negative	1	
Immunochromatography	Others	Negative	1	
Other	Abbott	Not Reported	1	
Other	Acon	Negative	1	
Other	Biomerieux	Negative	4	
Other	Siemens	Negative	1	
RIA	Not Reported	Negative	1	
RIA	Abon	Negative	2	

Copyright © 2021 [IACLD](#). All rights reserved.

[EQAP Panel Toggle navigation](#)

## • انجمن دکترای علوم آزمایشگاهی ایران

• دکتر علی عجمی

نوبل

• منوی اصلی

• صفحه نخست

• دوره های ارزیابی خارجی

## S1.2 Results

نمونه مجهول ایمنوسرولوژی برای آزمایش S1/2 HBsAg, HCVAb ,HIVAb

آزمایشگاه نوبل	EQAP-10814	1118 کد نمونه
----------------	------------	---------------

- [S1/2 HBs Ag](#)
- [S1/2 HCV Ab](#)
- [S1/2 HIV Ab](#)

### S1/2 HIV Ab

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ مثبت بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت مشخص شده است .

- ثبت نام کنندگان 1620
- شرکت کنندگان 1513
- کل درصد مشارکت 93.40 %
- پاسخهای دریافتی 1435
- درصد پاسخدهی 94.84 %

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	45	44	0	1	%97.78	%0.00	%2.22
ECL	123	123	0	0	%100.00	%0.00	%0.00
ELFA	112	109	0	3	%97.32	%0.00	%2.68
ELISA	1125	1073	7	45	%95.38	%0.62	%4.00
Immunochromatography	16	11	1	4	%68.75	%6.25	%25.00
Other	6	2	0	4	%33.33	%0.00	%66.67

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
CL	Abbott	Positive	19	←
CL	Architect	Positive	4	
CL	Biomerieux	Positive	2	
CL	Mindray	Not Reported	1	
CL	Mindray	Positive	8	
CL	Others	Positive	1	
CL	Roche	Positive	1	
CL	Siemens	Negative	1	
CL	Siemens	Positive	8	
ECL	Not Reported	Positive	2	
ECL	Abbott	Positive	9	
ECL	Acon	Positive	1	
ECL	Architect	Positive	2	
ECL	Biomerieux	Positive	5	
ECL	Cobas	Positive	5	
ECL	Mindray	Positive	3	
ECL	Roche	Positive	96	
ELFA	Not Reported	Positive	1	
ELFA	Abbott	Positive	1	
ELFA	Biomerieux	Negative	3	
ELFA	Biomerieux	Positive	106	
ELFA	Others	Positive	1	
ELISA	Not Reported	Not Reported	1	
ELISA	Not Reported	Negative	3	
ELISA	Not Reported	Positive	45	
ELISA	Abbott	Positive	1	
ELISA	Acon	Positive	1	
ELISA	Apdia	Positive	1	

Method	Kit	Result	Count	Your Answer
ELISA	Arya Mabna Tashkhis	Negative	1	
ELISA	Arya Mabna Tashkhis	Positive	64	
ELISA	Autobio Diagnostic	Positive	1	
ELISA	Bio-Pharm	Negative	1	
ELISA	Bio-Pharm	Positive	4	
ELISA	Biorad	Positive	4	
ELISA	Diapro	Positive	20	
ELISA	Euroimmun	Positive	2	
ELISA	Gbc	Not Reported	1	
ELISA	Gbc	Positive	38	
ELISA	General Biologicals(GB)	Not Reported	2	
ELISA	General Biologicals(GB)	Indeterminate	1	
ELISA	General Biologicals(GB)	Negative	6	
ELISA	General Biologicals(GB)	Positive	113	
ELISA	Mindray	Positive	2	
ELISA	Noyan Negin Parsian	Positive	2	
ELISA	Padtan Teb	Positive	7	
ELISA	Pars Azmoon	Positive	3	
ELISA	Pishtaz Teb	Not Reported	1	
ELISA	Pishtaz Teb	Indeterminate	6	
ELISA	Pishtaz Teb	Negative	32	
ELISA	Pishtaz Teb	Positive	625	
ELISA	Spinreact	Positive	1	
ELISA	Wantai	Not Reported	2	
ELISA	Wantai	Negative	2	
ELISA	Wantai	Positive	139	
Immunochromatography	Not Reported	Negative	1	
Immunochromatography	Not Reported	Positive	1	
Immunochromatography	Abon	Negative	3	

Method	Kit	Result	Count	Your Answer
Immunochromatography	Abon	Positive	6	
Immunochromatography	Addcare	Indeterminate	1	
Immunochromatography	Addcare	Positive	1	
Immunochromatography	Diagnostic	Positive	1	
Immunochromatography	General Biologicals(GB)	Positive	1	
Immunochromatography	Others	Positive	1	
Other	Not Reported	Negative	1	
Other	Acon	Negative	1	
Other	Biomerieux	Negative	2	
Other	Biomerieux	Positive	1	
Other	Siemens	Positive	1	
RIA	Abon	Positive	2	
RIA	Pishtaz Teb	Positive	1	

Copyright © 2021 [IACLD](#). All rights reserved.