

## انجمن دکترای علوم آزمایشگاهی ایران

• دکتر علی عجمی

• نوبل

• منوی اصلی

• [صفحه نخست](#)

• [دوره های ارزیابی خارجی](#)

• [خروج](#)

دوره چهل و دوم تابستان 1401

### S2 Results

نمونه مجهول ToRCH

آزمایشگاه نوبل	EQAP-10814	کد نمونه 1124
----------------	------------	---------------

• [Anti-CMV \(IgG\)](#)

• [Anti-HSV1 \(Anti-Herpes Type 1 IgG\)](#)

• [Anti-Rubella \(IgG\)](#)

• [Anti-Toxoplasma \(IgG\)](#)

### Anti-CMV (IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است .

• ثبت نام کنندگان 215

• شرکت کنندگان 195

• کل درصد مشارکت 90.70 %

• پاسخهای دریافتی 170

• درصد پاسخدهی 87.18 %

Method	Count			Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate

CL	61	59	0	2	%96.72	%0.00	%3.28
ECL	22	22	0	0	%100.0 0	%0.00	%0.00
ELFA	17	17	0	0	%100.0 0	%0.00	%0.00
ELISA	68	64	0	4	%94.12	%0.00	%5.88
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00
Other	2	2	0	0	%100.0 0	%0.00	%0.00

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
CL	Abbott	Positive	11	
CL	Architect	Positive	3	
CL	Autobio Diagnostic	Positive	4	
CL	Biomerieux	Positive	1	
CL	Dia Sorin	Negative	2	
CL	Dia Sorin	Positive	23	
CL	Diaplus	Positive	1	
CL	Liaison	Positive	5	
CL	Siemens	Positive	10	
ECL	Abbott	Positive	1	
ECL	Architect	Positive	1	
ECL	Biomerieux	Positive	1	
ECL	Cobas	Not Reported	1	
ECL	Roche	Positive	19	
ELFA	Biomerieux	Positive	16	
ELFA	Diesse-Tec	Positive	1	
ELISA	Not Reported	Positive	2	
ELISA	Chorus	Positive	1	
ELISA	Euroimmun	Positive	1	

Method	Kit	Result	Count	Your Answer
ELISA	Pishtaz Teb	Not Reported	2	
ELISA	Pishtaz Teb	Negative	4	
ELISA	Pishtaz Teb	Positive	54	←
ELISA	Vircell	Positive	6	
Other	Diesse-Tec	Positive	1	
Other	Pishtaz Teb	Positive	1	

Copyright © 2023 [IACLD](#). All rights reserved.

[EQAP Panel](#) Toggle navigation

## • انجمن دکترای علوم آزمایشگاهی ایران

- دکتر علی عجمی
- نوبل

- منوی اصلی
- [صفحه نخست](#)
- [دوره های ارزیابی خارجی](#)
- [خروج](#)

دوره چهل و دوم تابستان 1401

### S2 Results

نمونه مجهول ToRCH

آزمایشگاه نوبل	EQAP-10814	کد نمونه 1124
----------------	------------	---------------

- [Anti-CMV \(IgG\)](#)
- [Anti-HSV1 \(Anti-Herpes Type1 IgG\)](#)
- [Anti-Rubella \(IgG\)](#)
- [Anti-Toxoplasma \(IgG\)](#)

### Anti-HSV1 (Anti-Herpes Type1 IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است .

- ثبت نام کنندگان 215
- شرکت کنندگان 195
- کل درصد مشارکت 90.70 %
- پاسخهای دریافتی 78
- درصد پاسخدهی 40.00 %

Method	Total	Positive	Indeterminate	Count			Percent
				Negative	Positive	Indeterminate	
CL	12	12	0	0	%100.00	%0.00	%0.00
ECL	5	5	0	0	%100.00	%0.00	%0.00
ELFA	2	2	0	0	%100.00	%0.00	%0.00
ELISA	52	51	0	1	%98.08	%0.00	%1.92
Immunochromatography	1	1	0	0	%100.00	%0.00	%0.00
Other	6	6	0	0	%100.00	%0.00	%0.00

Method	Kit	Result	Count	Your Answer
CL	Not Reported	Positive	1	
CL	Autobio Diagnostic	Positive	1	←
CL	Dia Sorin	Positive	7	
CL	Liaison	Positive	3	
ECL	Abbott	Positive	1	
ECL	Cobas	Positive	1	
ECL	Pishtaz Teb	Positive	1	
ECL	Roche	Positive	2	
ELFA	Biomerieux	Not Reported	1	

Method	Kit	Result	Count	Your Answer
ELFA	Diesse-Tec	Positive	1	
ELFA	Orgentec	Positive	1	
ELISA	Not Reported	Not Reported	1	
ELISA	Not Reported	Positive	3	
ELISA	Architect	Positive	1	
ELISA	Autobio Diagnostic	Positive	1	
ELISA	Chorus	Positive	6	
ELISA	Euroimmun	Positive	2	
ELISA	ImmunoLab	Positive	1	
ELISA	Lion	Negative	1	
ELISA	Monobind	Not Reported	1	
ELISA	Orgentec	Positive	4	
ELISA	Pishtaz Teb	Not Reported	1	
ELISA	Pishtaz Teb	Positive	9	
ELISA	Vircell	Positive	24	
Immunochromatography	Chorus	Positive	1	
Other	Chorus	Positive	2	
Other	Diesse-Tec	Positive	2	
Other	Orgentec	Positive	2	

Copyright © 2023 [IACLD](#). All rights reserved.

[EQAP Panel](#) [Toggle navigation](#)

## • انجمن دکترای علوم آزمایشگاهی ایران

- دکتر علی عجمی
- نوبل

- منوی اصلی
- [صفحه نخست](#)
- [دوره های ارزیابی خارجی](#)
- [خروج](#)

دوره چهل و دوم تابستان 1401

## S2 Results

### نمونه مجهول ToRCH

کد نمونه 1124 EQAP-10814 آزمایشگاه نوبل

- [Anti-CMV \(IgG\)](#)
- [Anti-HSV1 \(Anti-Herpes Type1 IgG\)](#)
- [Anti-Rubella \(IgG\)](#)
- [Anti-Toxoplasma \(IgG\)](#)

### Anti-Rubella (IgG)

جدول شماره 1. توزیع فراوانی نتایج گزارش شده توسط شرکت کنندگان براساس روش تذکر. نتیجه مورد قبول برای این آزمایش پاسخ + بوده است. نتیجه ارائه شده توسط آزمایشگاه شما با علامت ← مشخص شده است .

- ثابت نام کنندگان 215
- شرکت کنندگان 195
- کل درصد مشارکت 90.70 %
- پاسخهای دریافتی 188
- درصد پاسخدهی 96.41 %

Method	Count				Percent		
	Total	Positive	Indeterminate	Negative	Positive	Indeterminate	Negative
CL	51	51	0	0	%100.00	%0.00	%0.00
ECL	24	24	0	0	%100.00	%0.00	%0.00
ELFA	17	17	0	0	%100.00	%0.00	%0.00
ELISA	94	90	2	2	%95.74	%2.13	%2.13
Immunochromatography	0	0	0	0	%0.00	%0.00	%0.00
Other	2	2	0	0	%100.00	%0.00	%0.00

Method	Kit	Result	Count	Your Answer
CL	Abbott	Positive	14	
CL	Architect	Positive	2	
CL	Autobio Diagnostic	Positive	5	
CL	Dia Sorin	Positive	18	
CL	Diaplus	Positive	1	
CL	Liaison	Positive	5	
CL	Others	Positive	1	
CL	Pishtaz Teb	Positive	1	
CL	Siemens	Positive	4	
ECL	Abbott	Positive	1	
ECL	Architect	Positive	2	
ECL	Cobas	Positive	1	
ECL	Dia Sorin	Positive	1	
ECL	Others	Positive	2	
ECL	Roche	Positive	17	
ELFA	Biomerieux	Positive	15	
ELFA	Diesse-Tec	Positive	1	
ELFA	Roche	Positive	1	
ELISA	Not Reported	Not Reported	1	
ELISA	Not Reported	Positive	3	
ELISA	Biomerieux	Positive	1	
ELISA	Chorus	Positive	2	
ELISA	Euroimmun	Positive	2	
ELISA	Pishtaz Teb	Indeterminate	2	
ELISA	Pishtaz Teb	Negative	2	
ELISA	Pishtaz Teb	Positive	74	←
ELISA	Vircell	Positive	8	
Other	Pishtaz Teb	Positive	2	